Product Datasheet

10X Citrate Buffer pH 8.0 NB900-66729

Unit Size: 500 ml

Store at room temperature.

www.novusbio.com



technical@novusbio.com

Protocols, Publications, Related Products, Reviews, Research Tools and Images at: www.novusbio.com/NB900-66729

Updated 10/23/2024 v.20.1

Earn rewards for product reviews and publications.

Submit a publication at www.novusbio.com/publications
Submit a review at www.novusbio.com/reviews/destination/NB900-66729



NB900-66729

10X Citrate Buffer pH 8.0

| 10X Citrate Buffer pH 8.0 | |
|-----------------------------|--|
| Product Information | |
| Unit Size | 500 ml |
| Concentration | Please see the protocols for proper use of this product. If no protocol is available, contact technical services for assistance. |
| Storage | Store at room temperature. |
| Preservative | 0.05% Sodium Azide |
| Buffer | Dilute one part buffer with nine parts distilled water. |
| Product Description | |
| Specificity/Sensitivity | 10x Citrate Buffer pH 8.0 for Heat Induced Epitope Recovery |
| Product Application Details | |
| Applications | Immunohistochemistry, Immunohistochemistry-Paraffin |
| Recommended Dilutions | Immunohistochemistry, Immunohistochemistry-Paraffin |
| Application Notes | The antigen retrieval protocol is recommended for use in tissues that have been fixed in formalin only. Ensure that the fixed sections are adequately embedded in paraffin. Cut tissue sections to 4-5 microns. Preparation of Working Solutions 1. The 10X concentrated format should be diluted tenfold with distilled or deionized water. 2. Mix one part of concentrated Antigen Retrieval Solution with nine parts of deionized or distilled water. 3. Shake the bottle vigorously to completely mix the components of the concentrate (the solution may separate into phases over time). 4. Store with cap tightly secured. Protocol Recommendations 1. Deparaffinize and rehydrate tissue sections. 2. Place slides into 1X retrieval solution in a slide container (e.g. Coplin jar, Tissue-Tek, staining dish or metal slide canister). 3. Retrieve sections under pressure. 4. After take-off reagent jar containing slides from pressure cooker, allow the slides to cool for 20 minutes to reach room temperature. 5. Wash slides in deionized water and then with wash buffer. Proceed with immunostaining recommendations in the antibody datasheet. 6. Gently rinse by gradually adding DI water to the solution, then remove slides and rinse with DI water. |





Novus Biologicals USA

10730 E. Briarwood Avenue Centennial, CO 80112 USA

Phone: 303.730.1950 Toll Free: 1.888.506.6887

Fax: 303.730.1966

nb-customerservice@bio-techne.com

Bio-Techne Ltd

19 Barton Lane Abingdon Science Park Abingdon, OX14 3NB, United Kingdom Phone: (44) (0) 1235 529449

Free Phone: 0800 37 34 15 Fax: (44) (0) 1235 533420 info.EMEA@bio-techne.com

Bio-Techne Canada

21 Canmotor Ave Toronto, ON M8Z 4E6 Canada

Phone: 905.827.6400 Toll Free: 855.668.8722 Fax: 905.827.6402

canada.inquires@bio-techne.com

General Contact Information

www.novusbio.com Technical Support: nb-technical@biotechne.com

Orders: nb-customerservice@bio-techne.com

General: novus@novusbio.com

Limitations

This product is for research use only and is not approved for use in humans or in clinical diagnosis. Support products are guaranteed for 6 months from date of receipt.

For more information on our 100% guarantee, please visit www.novusbio.com/guarantee

Earn gift cards/discounts by submitting a review: www.novusbio.com/reviews/submit/NB900-66729

Earn gift cards/discounts by submitting a publication using this product: www.novusbio.com/publications

